



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 15, 2005 INITIALS: DEC OWNER: WSP
DISTRICT: 3 ADDRESS: Highway 97 MP 56
LOCATION: Tippin # 42 COUNTY: Yakima
MAKE: Fahrer TYPE: Truck scale ACCURACY CLASS: III
MODEL: 42-9000-BW ID. NO.: G 608891 CAPACITY: 62500 SECTIONS: 2
MINIMUM GRADUATIONS: 50 lb REG. ELEMENT: dial PLATFORM: 8' x 12'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Center</u>	<u>4000</u>	<u>4000</u>	<u>0</u>				
	<u>8000</u>	<u>8000</u>	<u>0</u>	<u>Sec 1</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
	<u>12000</u>	<u>12000</u>	<u>0</u>				
	<u>16000</u>	<u>16000</u>	<u>0</u>	<u>Sec 2</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
<u>Center</u>	<u>20000</u>	<u>20,000</u>	<u>0</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 22205-7 and 8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by:

David Brown

Date:

June 16, 2005

State of Washington

County of

Thurston

Signed and sworn to (or affirmed) before
me on June 16, 2005 by

Michael J. Evans
(Signature of notary public)

Michael J. EVANS
Printed Name of Notary

My appointment expires

12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS